

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 3/14/2014

DBH
mmB

Instrument Serial Number:	010490	As of 14-Mar-14
---------------------------	--------	-----------------

Start Date	End Date	Agency Name
12-Mar-14		Chesapeake PD
06-Mar-14	12-Mar-14	DFS Central Lab
19-Aug-09	06-Mar-14	South Boston PD
27-May-08	19-Aug-09	DFS Central Lab

DBH
mmb

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010490 Worksheet Start Date 3/12/2014
Location Chesapeake Jail
Address 304 Albermarle Dr Chesapeake VA 23322
DFS Technician Donald Hall License No. 21250
☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 742 Reference Barometer (mm HG) 745
Reference Barometer(RB)Serial # 009109 RB Calibration Due 8/1/2014

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	
0.300	0.293	0.284	0.301	Sample 2	0.298
Precision		sample min	sample max	Sample 3	
0.001		0.297	0.298		0.297

Standard (sea level)	PA Target	minimum	maximum	Sample 1	
0.100	0.098	0.095	0.101	Sample 2	0.098
Precision		sample min	sample max	Sample 3	
0.001		0.098	0.099		0.099

Dry gas standard Lot No. (with tank no.) AG400202-37

☐ Replaced dry gas standard (+O-ring)

☒ Installed at Location

☐ Removed to DFS-Central

Supplies

Mouthpieces 300

Certificates of Analysis 200

Operator Worksheet 150

Other:

Notes:

Instrument Serial Number

010490

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B Hall

Date

3/14/14

Issuing Analyst

Michelle Bullock

Date

3/17/14WBL

D6H
mmB




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 03/12/2014
TEST EQUIPMENT NUMBER 010490		

RESULTS: TIME SAMPLE TAKEN 14:31 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 12 Mar 2014 at 16:53

DBH
mmB

Test Results

Instrument Serial Number 010490

Test # 000589 Subject Test

Test Location 1 Department of
Test Date 12 Mar 2014

Test Location 2 Forensic Science
Test Time 14:24
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 12 Mar 2014 End Time 14:32

Result Time 14:31

Result Date 12 Mar 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 14:25

Data Type BLK

Sample Value 0.000

Sample Time 14:25

Data Type CHK

Sample Value 0.098

Sample Time 14:26

Data Type BLK

Sample Value 0.000

Sample Time 14:28

Data Type SUBJ

Sample Value 0.000

Sample Time 14:28

Data Type BLK

Sample Value 0.000

Sample Time 14:29

Data Type SUBJ

Sample Value 0.000

Sample Time 14:31

Data Type BLK

Sample Value 0.000

Sample Time 14:32

Standard Type Dry Gas Std

Standard Value 0.098

Standard Lot Number AG400202-37

Standard Expiration Date 01/02/2016

Tank Pressure 1120

Barometric Pressure 741 mmHg

Blow Sample Number 1 Blow Duration 3.30 sec

Blow Volume 1824 cc End-of-Blow Time 14:28

Blow Sample Number 2 Blow Duration 3.83 sec

Blow Volume 1864 cc End-of-Blow Time 14:31

Tamper Evident Stamp 4a9fa04a

Test Status Code 0

Test Status Success

Intox EC/IR-II: Accuracy Check

DBH
mmB

Department of Forensic Science

Serial Number: 010490 Test Number: 587

Test Date: 03/12/2014 Test Time: 14:06 EDT

Dry Gas Target: 0.293

Lot Number: AG322702-06 Exp Date: 08/15/2015

Tank Pressure: 301 psi Barometric Pressure: 741 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	14:07
CHK	0.298	14:07
BLK	0.000	14:09
CHK	0.298	14:10
BLK	0.000	14:12
CHK	0.297	14:13

Test Status: *Success*

Calibration CRC: D62CC7E5

Intox EC/IR-II: Accuracy Check

DBH
mmB

Department of Forensic Science

Serial Number: 010490 Test Number: 588

Test Date: 03/12/2014 Test Time: 14:16 EDT

Dry Gas Target: 0.098

Lot Number: AG400202-37 Exp Date: 01/02/2016

Tank Pressure: 1120 psi Barometric Pressure: 741 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	14:16
CHK	0.098	14:17
BLK	0.000	14:19
CHK	0.099	14:20
BLK	0.000	14:21
CHK	0.098	14:22

Test Status: *Success*

Calibration CRC: D62CC7E5